

PRODUCT IDENTIFICATION	
Product name	Model/number
PolyMem Non-Adhesive Dressings	REF 5022; REF 5033; REF 5044; REF 5055 REF 5077; REF 5124; REF 5244; REF 5824
PolyMem Adhesive Film Dressings	REF 203; REF 0230; REF 405; REF 606 REF 3412
PolyMem Cloth-Backed Adhesive Dressings	REF 7203; REF 7405; REF 7606; REF 7031 REF 7042; REF 7048; REF 7412
PolyMem MAX	REF 5035; REF 5045; REF 5050; REF 5088
Shapes by PolyMem	REF 3709; REF 8086; REF 8053; REF 8023
PolyMem Tube Dressings	REF 5335; REF 5333
PolyMem WIC Cavity Fillers	REF 5712; REF 5731; REF 5733; REF 5750 REF 5788
PolyMem Finger/Toe Dressings	REF 4401; REF 4402; REF 4403; REF 4404 REF 4405
PolyMem Surgical Dressings	REF 0481; REF 0511; REF 0512
OptiMem by PolyMem Non-Adhesive Dressings	REF 0707; REF 1010; REF 1719
PolyMem Sample Packs	REF 3000

MANUFACTURER		
Name of company	Address	Management Representative
Ferris Mfg. Corp.	5133 Northeast Parkway Fort Worth, TX 76106 U.S.A.	Paul Kay

AUTHORIZED REPRESENTATIVE		
Name of company	Address	Telephone/email
Emergo Europe	Prinsessegracht 20	+31.70.345.8570 - phone
	2514 AP The Hague	+31.70.346.7299 - fax
	The Netherlands	service@emergogroup.com

REGISTRATION INFORMATION	
Notified Body and ID #	Date CE Marking first applied
TÜV Rheinland LGA Products GmbH	Original Certificate Issued 02/10/1997
Tillystraße 2	CE Marking first Applied to PolyMem:
90431 Nürnberg	02/10/1997
Germany	
ID# 0197	

CONFORMITY ASSESSMENT		
Device classification	Route to compliance	
Class IIb	Annex II (excluding Section 4) of MDD 93/42/EEC	
Rule 4		

**Ferris Mfg. Corp.** declares that the above mentioned products meet the provision of the MDD 93/42/EEC for Medical Devices. This Declaration of Conformity is valid in conjunction with the applicable Product Release Form for the lot.

**COMPANY REPRESENTATIVE:** Paul Kay

TITLE: Quality System Management Representative

SIGNATURE:

JUN 2 0 2017

900CUIVIENT CUNTROL

D.O.C. -1 REV-20 CR# 17-326



PRODUCT IDENTIFICATION	
Product name	Model/number
PolyMem Non-Adhesive Silver Dressing	REF 1022; REF 1044; REF 1047;
(Various Sizes)	REF 1077; REF 1124; REF 1244;
	REF 1824
PolyMem Cloth-Backed Adhesive Silver Dressing	REF 1766
PolyMem MAX Silver by PolyMem	REF 1088; REF 1045
Shapes Silver by PolyMem	REF 1709; REF 1886; REF 1853;
	REF 1823
PolyMem WIC Silver Cavity Filler	REF 1331; REF 1333; REF 1788
PolyMem WIC Silver Rope	REF 1814
PolyMem Silver Finger/Toe Dressings	REF 1401; REF 1402; REF 1403;
-	REF 1404; REF 1405
PolyMem Silver Surgical Dressings	REF 1511; REF 1512

MANUFACTURER		
Name of company	Address	Representative
Ferris Mfg. Corp.	5133 Northeast Parkway Fort Worth, TX 76106	Paul Kay
	U.S.A.	

AUTHORIZED REPRESENTATIVE		
Name of company	Address	Telephone/email
Emergo Europe	Prinsessegracht 20	+31.70.345.8570 - phone
	2514 AP The Hague	+31.70.346.7299 - fax
	The Netherlands	service@emergogroup.com

REGISTRATION INFORMATION		
Notified Body and ID #	Date CE Marking first applied	
TÜV Rheinland LGA Products GmbH Tillystraße 2 90431 Nürnberg Germany ID# 0197	Original Certificate Issued 02/10/1997 CE Marking first Applied to PolyMem Silver : 2004	

CONFORMITY ASSESSMENT	
Device classification	Route to compliance
Class IIb	Annex II (excluding Section 4) of MDD 93/42/EEC
Rule 4	

Ferris Mfg. Corp. declares that the above mentioned products meet the provision of the MDD 93/42/EEC for Medical Devices. This Declaration of Conformity is valid in conjunction with the splicable Product Release Form for the lot.

**COMPANY REPRESENTATIVE:** Paul Kay

TITLE: Quality System Management Representative

\_ DATE: 2017 Jume 19

JUN 20 2017

DOCUMENT CUNTROL



PRODUCT IDENTIFICATION		
Product name	Model/number	
Nursicare by PolyMem®	REF 400	
Sterile Breast Pad	available as	
	REF 400-AR; REF 400-BN; REF 400-BR; REF 400-CA; REF 400-CH;	
	REF 400-DE; REF 400-EN; REF 400-ES; REF 400-FI; REF 400-IL;	
	REF 400-NO; REF 400-TW; REF 400-US; REF 400S	

MANUFACTURER		
Name of company	Address	Management Representative
Ferris Mfg. Corp.	5133 Northeast Parkway Fort Worth, TX 76106 U.S.A.	Paul Kay

AUTHORIZED REPRESENTATIVE		
Name of company	Address	Telephone/email
Emergo Europe	Prinsessegracht 20	+31.70.345.8570 - phone
	2514 AP The Hague	+31.70.346.7299 - fax
	The Netherlands	service@emergogroup.com

REGISTRATION INFORMATION	
Notified Body and ID #	Date CE Marking first applied
TÜV Rheinland LGA Products GmbH Tillystraße 2 90431 Nürnberg Germany ID# 0197	Original Certificate Issued 02/10/1997 CE Marking first Applied to Nursicare: April 15, 2013

CONFORMITY ASSESSMENT	
Device classification	Route to compliance
Class I (s)	Annex II (excluding Section 4) of MDD 93/42/EEC
Rule 4 (sterile)	

**Ferris Mfg. Corp.** declares that the above mentioned products meet the provision of the MDD 93/42/EEC for Medical Devices. This Declaration of Conformity is valid in conjunction with the applicable Product Release Form for the lot.

**COMPANY REPRESENTATIVE: Paul Kay** 

**TITLE:** Quality System Management Representative

SIGNATURE:

DATE:

ISSHED

JUN 20 2017

DOCUMENT CONTROL



PRODUCT IDENTIFICATION		
Product name	Model/number	
SportsWrap™	REF 9316 ; REF 9424; REF 9524	
(With Fabric OverWrap)		
SportsWrap™ Roll	REF 9336; REF 9536; REF 9548	
SportsWrap™ ST Roll	REF 9120	

MANUFACTURER		
Name of company	Address	Management Representative
Ferris Mfg. Corp.	5133 Northeast Parkway Fort Worth, TX 76106	Paul Kay
	U.S.A.	

AUTHORIZED REPRESENTATIVE		
Name of company	Address	Telephone/email
Emergo Europe	Prinsessegracht 20	+31.70.345.8570 - phone
	2514 AP The Hague	+31.70.346.7299 - fax
	The Netherlands	service@emergogroup.com

REGISTRATION INFORMATION	
Notified Body and ID #	Date CE Marking first applied
N/A	Original Certificate Issued 02/10/1997
Self Declaration	CE Marking first Applied to SportsWrap: April 2007

CONFORMITY ASSESSMENT	
Device classification	Route to compliance
Class I	Annex VII of MDD 93/42/EEC
Rule 1	

**Ferris Mfg. Corp.** declares that the above mentioned products meet the provision of the MDD 93/42/EEC for Medical Devices. This Declaration of Conformity is valid in conjunction with the applicable Product Release Form for the lot.

**COMPANY REPRESENTATIVE: Paul Kay** 

TITLE: Quality System Management Representative

SIGNATURE:

DATE: 2017 Jone 19

JUN 20 2017 DOCUMENT CONTROL