

## Declaration of Conformity



**Manufacturer:** Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, Hi-tech Industrial Park,  
Nanshan, Shenzhen, 518057, P. R. China

**EC-Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80  
20537 Hamburg, Germany

**Product:** Electrocardiograph (Including Accessories)

**Model:** BeneHeart R12、 BeneHeart R12A

**Classification:** IIa (According to Rule 10 of MDD Annex IX)

**Conformity Assessment Route:** MDD Annex II excluding(4)

**We herewith declare that the above mentioned products meet the provisions of the Council Directive 93/42/EEC concerning Medical Device, as amended by 2007/47/EC. All supporting documentations are retained under the premises of the manufacturer.**

### Standards Applied:

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

**Notified Body:** TÜV SÜD Product Service GmbH  
Ridlerstraße 65  
80339 München, Germany.

**Notified Body No. :** 0123

**Start of CE-Marking:**

**Place, Date of Issue:** Shenzhen, 2014.1.8.

**Signature:** 

**Name of Authorized Signatory:** Mr. Guan Zhiyong

**Position Held in Company:** Director Technical Regulation

## Applied Standards List

**Product:**            **Electrocardiograph**

**Model:**             **BeneHeart R12、 BeneHeart R12A**

### Standards Applied:

<b>EN ISO 14971: 2012</b>	Medical devices – Application of risk management to medical devices
<b>EN 1041: 2008</b>	Information supplied by the manufacturer of medical devices
<b>EN ISO 15223-1: 2012</b>	Medical devices-Symbols to be used with medical device labels, labeling and information to be supplied-Part 1: General Requirements
<b>EN ISO 10993-1: 2009/AC:2010</b>	Biological evaluation of medical devices - Part 1: Evaluation and testing
<b>EN 60601-1: 1990+A1:1993+A2:1995</b>	Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance
<b>EN 60601-1-2: 2007/AC:2010</b>	Medical Electrical Equipment – Part 1-2: General Requirements for Basic Safety and Essential Performance- Collateral Standard: Electromagnetic compatibility - Requirements and tests
<b>EN 60601-1-4: 1996+A1:1999</b>	Medical electrical equipment - Part 1-4: General requirements for safety - Collateral standard: Programmable electrical medical systems
<b>EN 60601-1-6: 2007/AC:2010</b>	Medical electrical equipment - Part 1-6: General Requirements for basic safety and essential performance -collateral standard: usability
<b>EN 60601-2-25:1995+A1:1999</b>	Medical electrical equipment – Part 2-25: Particular requirements for the safety of electrocardiographs
<b>EN 60601-2-51: 2003</b>	Medical electrical equipment – Part 2-51: Particular requirements for safety, including essential performance, of recording and analysing single channel and multichannel electrocardiographs
<b>EN 62366:2008</b>	Medical devices -- Application of usability engineering to medical devices
<b>EN 62304:2006 /AC:2008</b>	Medical device software – Software life cycle processes
<b>ANSI/AAMI EC11: 1991/(R)2007</b>	Diagnostic electrocardiographic devices