



DECLARATION OF CONFORMITY

Legal Manufacturer:	Becton, Dickinson and Company Limited, Donore Road, Drogheda, Co. Louth, Ireland.
Authorised Representative:	NA
Manufacturing Site:	Manufactured on behalf of Becton, Dickinson and Company, Limited by: Becton Dickinson S.A. Crta. Mequinenza s/n. 22520 – Fraga HUESCA – SPAIN
Products:	BD Microlance™ 3 Needles (see attached list for the different types)
Classification:	Class IIa, Rule 6
Conformity Assessment Route:	Annex VII and Annex V
GMDN:	59230; sterile 59229; non-sterile


We herewith declare that for the above mentioned products, their product design meets the provisions of the European Medical Device Directive 93/42/EEC of 14 June 1993 and Irish Regulation; S.I. No. 252 of 1994 (as amended by Regulation 5 of S.I. 110/2009) concerning medical devices. All supporting documentation is retained at the premises of the manufacturer or subcontractor.

Harmonized standards:	EN 556-1:2001/AC:2006, EN 980:2008, EN 1041:2008/A1:2013, EN1707:1996, EN 20594-1:1993/AC:2009, EN ISO 10993-1:2009/AC:2010, EN ISO 10993-3:2014, EN ISO 10993-4:2009, EN ISO 10993-5:2009, EN ISO 10993-7:2008/AC2009, EN ISO 10993-11:2009, EN ISO 10993-12:2012, EN ISO 10993-13:2010, EN ISO 10993-15:2009, EN ISO 10993-17:2009; EN ISO 10993-18:2009; EN ISO 11135-1:2007, EN ISO 11138-2:2009, EN ISO 11737-1:2006/AC:2009, EN ISO 11737-2:2009, EN ISO 11607-1:2009, EN ISO 11607-2:2006, EN ISO 13485:2012, ISO13485:2003, EN 14155:2011, EN ISO 14971:2012,
Non -Harmonised Standards	ISO 14644-1:1999, ISO 10993-2:2006, ISO 10993-10:2009, ISO 9626: 1995, ISO 2859-1:1999, ISO 7864-1:1996; ISO15223-1:2012, ISO 594-1:1993, ISO 594-2:1998, ISO 6009; 1995/COR 2008
Notified Body:	NSAI 1 Swift Square, Northwood, Santry Dublin 9, Ireland Telephone: 353-1-807-3800 Number of Notified Body: 0050
CE Certificate number:	252.157
Date of issue of original CE certificate:	24 July 2012

Date: 23 November 2017

Date: 23 November 2017

Name Andrew Roche
Function QA/RA
Business Unit Becton, Dickinson and Company Limited


Cormac Reynolds
Director
Becton, Dickinson and Company Limited

DECLARATION OF CONFORMITY

BD MICROLANCE NEEDLES

STERILE NEEDLES

CATALOG No.	PRODUCT DESCRIPTION	
	(Imperial, inch)	(S.I.) mm
300094	22G x 2	0.7 x 50
300300	26G x 3/8	0.45 x 10
300400	25G x 1	0.5 x 25
300600	25G x 5/8	0.5 x 16
300635	27G x 1/2	0.4 x 13
300637	16G x 1 1/2	1.6 x 40
301155	21G x 2	0.8 x 50
301500	19G x 1 1/2	1.1 x 40
301700	19G x 1	1.1 x 25
301750	19G x 2	1.1 x 50
301900	18G x 2	1.2 x 50
302200	27G x 3/4	0.4 x 19
303800	26G x 1/2	0.45 x 13
304000	30G x 1/2	0.3x 13
302809*	30G x 1/2	0.3x 13
304100	24G x 1	0.55 x 25
304300	26G x 5/8	0.45 x 16
304622	18G x 1 1/2	1.2 x 40

***Pharmaceutical Quality**

DECLARATION OF CONFORMITY

BD MICROLANCE NEEDLES

NON-STERILE NEEDLES (BULK)

CATALOG No.	PRODUCT DESCRIPTION	
	(Imperial, inch)	(S.I.) mm
300206	25G x 5/8"	0.5 x 16
300284	21G x 5/8	0.8 x 16
300403	25G x 1	0.5 x 25
300801	23G x 1	0.6 x 25
300802	21G x 1 ½	0.8 x 40
300806	21Gx 1	0.8 x 25
300807	20Gx 1	0.9 x 25
300808	20G X 1 ½	0.9 x 40
300901	22G X 1 ¼	0.7 x 30
300932	30G X ½	0.3 x 13
301296	27G X ½	0.4 x 13
301530	19G X 1½	1.1 x 40
302424	14G X 1 ¼	2.1 x 30
303803	26G X ½	0.45 x 13
304623	26G X 3/8	0.45 x 10
304627	23G X 1	0.6 x 25
304630	21Gx2	0.8 x 50
304636	18G X 2	1.1 x 50
304637	27G X ¾	0.4 x 19
304641	24G X 1	0.55 x 25
304644	18G X 1½ SB TW	1.2 x 40
304728	22G X 1	0.7 x 25
304729	22G X 1½	0.7 x 40
305146	23G X 1 ¼	0.6 x 30
306030	25G x 5/8	0.5 x 16